
# **Quick Guide**






### fasciotens®Hernia at a glance


## Preparation of the patient



Stitch the sutures (polyfil USP 2 > 50 cm) as a U-suture (2-3 cm wide) with a distance of 1 cm to the fascial edge.




Protect the abdominal organs with moist abdominal drapes during the traction phase.



Option: For easier handling, you can cross the sutures before tensioning.

### Assembling fasciotens®Carrier







Video guide

Just scan the QR code and watch the clip.

# Attaching fasciotens<sup>®</sup>Hernia to the fasciotens<sup>®</sup>Carrier





#### Tightening the traction sutures under pre-tension



Press down the screw head to create a pre-tension of 14 kg. Hold the tension until ALL the individual sutures are tensioned and then slowly release the pre-tension.

The individual sutures should be retightened every 2 min. If the individual sutures cannot be tightened any more, the tensile force can be increased further by turning the screw head.





Always clamp the sutures in the frame from the outside to the inside.

Please note the instructions for use!





**fasciotens GmbH** Moltkeplatz 1 D-45138 Essen Germany **www.fasciotens.de** info@fasciotens.de +49 (0)221 17738500

#### Support:

support@fasciotens.de Tel.: +49 (0)221 17738500

#### fasciotens GmbH

Aachener Str. 1053-1055 D-50858 Cologne Germany